KEYSURGICAL®

Key® Scope Cleaning Brushes
BR-3300/BR-3305/BR-3307

INTENDED USE:

These brushes are intended to be used as an aid in cleaning and disinfecting flexible endoscopes. The BR-3300 & BR-3305 are used to scrub the inside of the endoscope channel and loosen debris present. The BR-3307 is used to scrub the inside of the suction well of the endoscope. Do not use these instruments for any purpose other than their intended use.

PREPARATION, INSPECTION, & USE:

1. Do not use reusable cleaning brushes for cytological tissue sampling or any other diagnostic or therapeutic purposes. Patient injury, cross-contamination, and equipment damage may occur.
2. The brush should be inspected between uses and replaced when worn, frayed, bent, or otherwise damaged. **This inspection is the responsibility of the user.** Worn bristles are ineffective in cleaning, and damaged brushes may damage the endoscope.
3. Ensure brush is the proper size for the channel it is being used to clean.
4. BR-3300 & BR-3305: Advance the brush in short strokes to avoid kinking. Do not rotate the brush inside the channel of the endoscope. Scope Cleaning Brushes shall not be coiled in loops smaller than 10" (25cm). Refer to a particular endoscope user manual for instructions on how to clean the endoscope channel.
5. BR-3307: Read the user manual for the endoscope and perform the necessary preparation and inspection. Refer to the particular endoscope user manual for instructions on how to clean the suction well.

REPROCCESSING:

1. Patient debris and reprocessing chemicals are hazardous. Wear protective equipment.
2. Thoroughly clean the brush to remove adhering debris. If the brush is not thoroughly cleaned, residual organic debris may solidify and it may be difficult to properly disinfect the brush.
3. When performing ultrasonic cleaning and disinfection, do not coil the Scope Cleaning Brush into a diameter smaller than 10" (25cm). The brush can be damaged if coiled too tightly.
4. Disinfect cleaning brush in High-Level Disinfectant according to the HLD manufacturer's instructions. Rinse brush thoroughly with sterile water. Allow to dry for storage.
5. Only trained and qualified personnel should be given the responsibility to reprocess cleaning brushes.

STORAGE:

Store cleaning brush at room temperature in a dry, clean location away from direct sunlight.

www.keysurgical.com
BROC-3077, Rev C, 3 April 2012
T: 800.541.7995